

# The TF20 microscope interfaces

## The microscope console

**Location:** right in front of you on the desk

**Role:** It provides direct control of some of the most important microscope parts:

- Magnification
  - Electron dose: spot size (C1 lens) and beam spread (C2 lens)
  - Focus (objective lens current)
  - Stage position (with the joystick and Z buttons)
  - Beam position (with the trackball)
  - and many other actions through the programmable switches (L/R1-2-3) and Multifunction X/Y knobs.
- 

## Gatan DigitalMicrograph User Interface (DM)

**Location:** On the left screen

**Role:** It controls the US10001 Camera.

- the user will use it to :
    - prepare Dark and gain references.
    - during screening, to acquire images in continuous mode.
- 

## SerialEM User Interface (SEM)

**Location:** On the right screen

**Role:** It provides camera/microscope commands to allow the user to screen its grids and setup an automated data collection.

- **The beginner user will use it to:**
    - record the Atlases.
    - select squares and prepare the square maps at eucentric height.
    - select targets (position).
    - screen targets at higher magnification (as single image or montage).
    - center the different beam modes all together on the camera axis.
  - **The expert user will use it to:**
    - perform all the actions described above.
-

- perform beam corrections (astigmatism, coma-free).
- perform calibrations.

---

## The TEM User Interface (TUI)

**Location:** On left screen

**Role:** It controls all the microscope systems.

\* It is divided into 4 zones:

- Zone1: TUI tabbed pages hosting different commands allowing to control the ME and to know its working state

\*Zone1:TUI tabbed . \*Zone 2 : Window giving values of working state : HT value, stage position, magnification, etc ... \*Zone 3 : Additional TUI Tabbed page \*Zone4 : Window defining functions of control panel button:□

- The beginner user will use it to:
  - load/unload the grids in the column.
  - change and center the Condenser 2 and objective apertures.
  - control the turbo pump.
  - perform beam-shift adjustments.
  - manually set eucentric height.
- The expert user will use it to:
  - perform all the actions described above.
  - align the microscope.
  - troubleshoot the system.
  - perform cryo-cycle of the column/autoloader.

---

### The FluCam User Interface (FC)

Location: On the right screen

Role: It allows the visualization of the electron beam and the sample image.

The beginner user will use it to:

The expert user will use it to:

From:

<https://bsi.inscog.eu/> - **BSI wiki**

Permanent link:

[https://bsi.inscog.eu/doku.php?id=tf20\\_interface&rev=1679502989](https://bsi.inscog.eu/doku.php?id=tf20_interface&rev=1679502989)

Last update: **2023/11/01 20:15**

